

GEOSTATS PTY LTD

Mining Industry Consultants
Reference Material Manufacture and Sales

Certified Base Metal Reference Material Product Code

GBM307-5


Certified Control Values

Base Metal Analyses

Element	Grade	Standard Deviation	No of Analyses	Confidence Interval
Nickel (ppm)	29	7	50	+/- 2
Copper (ppm)	31	5	56	+/- 1.4
Zinc (ppm)	18	5	53	+/- 1.5
Lead (ppm)	34	10	53	+/- 2.9
Arsenic (ppm)	23	13	37	+/- 4.2
Cobalt (ppm)	4	1	25	+/- 0.6
Silver (ppm)	1.0	0.7	17	+/- 0.4

CRM Details

Control Statistic Details	Neutron Activation Analysis Results (ppm)		Major Elements Fusion / XRF (%)	
	<p>Control statistics were produced from results accumulated in the : <u>April-2007</u> Geostats Pty Ltd Laboratory Round Robin Program. <u>17</u> laboratories (at least) tested this material for base metal content.</p> <p>Source Material Prior to homogenisation and testing, this material was sourced from Laterite Material ferruginous</p> <p>Colour Designation Moderate brown</p> <p>Usage This product is for use in the mining industry as reference materials for monitoring and testing the accuracy of laboratory assaying.</p> <p>Preparation and Packaging All standards are dried in an oven for a minimum of 12 hours at 110C. The dry material is then pulverised to better than 75 micron (nominal mean of 45 micron) using an Air Classifier. The material is then homogenised and stored in a sealed, stable container ready for final packaging.</p> <p>Materials are statistically sampled from stores, then packaged into either heat sealed, air tight, plastic pulp packets or screw top sealed plastic containers ready for distribution. All packaging has been chosen to ensure minimal contamination from outside sources during shipment, use and storage.</p> <p>Assay Testwork All standards are tested thoroughly in the Geostats bi-annual laboratory survey. This involves assaying by a minimum of 50 reputable laboratories selected from across the world using a variety of methods (including AR, 3AD, 4AD and ICP, AAS and XRF). Results are compiled into a comprehensive report detailing statistics for each standard. Assay distributions are checked and processed statistically, producing monitoring statistics for these standards. Materials are tested regularly to ensure stability and homogeneity.</p>	<p>Antimony 0.5</p> <p>Arsenic 29</p> <p>Barium <50</p> <p>Bromine 3.1</p> <p>Cadmium <5</p> <p>Cerium 25</p> <p>Caesium 0.9</p> <p>Chromium 220</p> <p>Cobalt 4</p> <p>Europium <1</p> <p>Gold ppb 59</p> <p>Hafnium 12</p> <p>Iridium ppb <50</p> <p>Iron % 20</p> <p>Lanthanum 9</p> <p>Lutetium <0.2</p> <p>Molybdenum 7</p> <p>Nickel 34</p> <p>Rubidium <5</p> <p>Samarium 1.4</p> <p>Scandium 11</p> <p>Selenium <5</p> <p>Sodium % 0.11</p> <p>Tantalum 2.2</p> <p>Tellurium <10</p> <p>Terbium <0.5</p> <p>Thorium 67</p> <p>Tin <100</p> <p>Tungsten 3</p> <p>Uranium 6.4</p> <p>Ytterbium <2</p> <p>Zinc <100</p> <p>Zirconium 360</p> <p>Calcium% nr</p> <p>Potassium % nr</p> <p>Silver <1</p> <p>Mercury nr</p> <p>Neodymium nr</p> <p>Strontium nr</p>	<p>Fe nr</p> <p>SiO2 nr</p> <p>Al2O3 nr</p> <p>TiO2 nr</p> <p>MnO nr</p> <p>CaO nr</p> <p>P nr</p> <p>S nr</p> <p>MgO nr</p> <p>K2O nr</p> <p>Na2O nr</p> <p>LOI1000 nr</p>	



GBM307-5

Geostats Pty Ltd, Certified Base Metal Reference Material, Product Code :

10A Marsh Close, O'Connor, Western Australia 6163

Phone : +61 8 9314 2566, Fax : +61 8 9314 3699

e-mail : pjh@geostats.com.au, srr@geostats.com.au

Website <http://www.geostats.com.au>