

GEOSTATS PTY LTD

Mining Industry Consultants
Reference Material Manufacture and Sales

Certified Pulp Rare Earth Reference Material

GRE-10

Certified Control Values

Method	Element	Units	Grade	Standard Deviation	No of Analyses	95% Confidence Interval
Fusion / ICP	Al	%	7.52	0.16	47	+/- 0.05
	Ba	ppm	8943	366	50	+/- 105
	Ca	%	1.5	0.1	49	+/- 0.1
	Ce	ppm	4064	217	40	+/- 71
	Dy	ppm	58	2	50	+/- 1
	Er	ppm	9.4	0.9	50	+/- 0.3
	Eu	ppm	57.7	2.0	50	+/- 0.6
	Fe	%	17.25	0.62	50	+/- 0.18
	Gd	ppm	150	7	50	+/- 3
	Ho	ppm	6.2	0.2	50	+/- 0.1
	La	ppm	834	31	50	+/- 9
	Lu	ppm	0.4	0.1	30	+/- 0.1
	Mg	%	0.51	0.01	50	+/- 0.01
	Mn	ppm	4951	145	47	+/- 43
	Nb	ppm	93	17	50	+/- 5
	Nd	ppm	2914	130	50	+/- 38
	P	ppm	8703	329	40	+/- 107
	Pr	ppm	721	15	40	+/- 6
	S	%	0.04			
	Sc	ppm	33	5	40	+/- 2
	Si	%	20.3	0.6	50	+/- 0.2
	Sm	ppm	283	10	50	+/- 3
	Sr	ppm	723	23	50	+/- 7
	Ta	ppm	1	0	30	+/- 1
	Tb	ppm	16.0	0.8	50	+/- 0.3
	Th	ppm	340	12	50	+/- 4
	Tm	ppm	0.8	0.1	40	+/- 0.1
	U	ppm	24.9	1.0	50	+/- 0.3
	Y	ppm	137	6	50	+/- 2
	Yb	ppm	3.8	0.2	49	+/- 0.1
	Zr	ppm	82	23	38	+/- 8

CRM Details

Control Statistic Details

Control values for this material were determined during a dedicated certification program.

Certification Date

This material was certified with the above values on the 30th of May 2023.

Source Material

Prior to homogenisation and testing, this material was sourced from:
Gascoyne region, Western Australia

Material Type

Pulp Rare Earth ore, 10g samples.

Usage

This product is for use in the mining industry as reference materials for monitoring and testing the accuracy of laboratory assaying.

Preparation and Packaging

This reference material was dried in an oven for a minimum of 8 hours at 110C. The dry material is then pulverised in a bowl and puck mill and homogenised in a vee-blender. The material is then stored in a sealed, stable container ready for final packaging.

Materials are statistically sampled from stores, then packaged into heat sealed, air tight, plastic packets ready for distribution. All packaging has been chosen to ensure minimal contamination from outside sources during shipment, use and storage.

Assay Testwork


This standard was tested in a dedicated certification program. In late 2022, 10 x 20g samples were sent to 6 laboratories for fusion / ICP analyses. Assay distributions are checked and processed statistically, producing monitoring statistics for these standards. Materials are tested regularly to ensure stability and homogeneity.

Stability

This product remains stable in its original packaging, away from direct sunlight.

Material Safety

This product is not hazardous and non-toxic.



20 Hines Road, O'Connor, Western Australia, 6163

Phone : +61 8 9314 2566

e-mail : info@geostats.com.au

Website : <http://www.geostats.com.au>

Geostats Pty Ltd, Certified Rare Earth Reference Material, Product Code:

GRE-10